

**Notice of Allowability**

Application No.

10/632,507

Examiner

Stanley J. Pruchnic, Jr.

Applicant(s)

CARTER ET AL.

Art Unit

2859

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 5 October 2004.
2. ☒ The allowed claim(s) is/are 1.
3. ☒ The drawings filed on 5 October 2004 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☐ All    b) ☐ Some\*    c) ☐ None    of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  6. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
    - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
      - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
    - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- |   |  |
|---|--|
| 1. <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)   | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)            |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                | 6. <input type="checkbox"/> Interview Summary (PTO-413),<br>Paper No./Mail Date _____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),<br>Paper No./Mail Date _____ | 7. <input type="checkbox"/> Examiner's Amendment/Comment                               |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit<br>of Biological Material          | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance   |
|   | 9. <input type="checkbox"/> Other _____.   |

### ***Reasons for Allowance***

1. The following is an examiner's statement of reasons for allowance: Applicant's arguments are persuasive with respect to the amended claim. Although the use of a single NiCu alloy, Monel, to calibrate a TGA is generally known, the use of several such alloys over the claimed limited range for the amount of Cu is critical to the invention which is needed for calibration of a vacuum thermogravimetric analyzer (VTGA), as described in the Specification on Pages 5 and 15. Moreover, none of the prior art discloses a TGA or VTGA wherein the set-point temperature is offset higher than the known Curie temperature of each slug as claimed by Applicant.

Claim 1 is allowable because the prior art fails to teach or fairly suggest a method to provide a temperature calibration for a vacuum thermogravimetric analyzer (VTGA) comprising:

selecting a plurality of ferromagnetic slugs wherein each slug includes an alloy including Ni and Cu, and wherein the amount of Cu is within the range of 15% to 28%;  
and

setting a temperature of the VTGA to a temperature corresponding to a set-point temperature greater than the Curie temperature of each slug by an amount equal to a first offset value as recited in claim 1 in combination with the other limitations of the claim.

2. Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

### ***Conclusion***

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. The prior art cited in PTO-892 and not mentioned above disclose related temperature calibration methods for thermogravimetric analyzer systems including the use of a plurality of ferromagnetic samples having magnetic transitions:

Art Unit: 2859

- Norem, S. D.; O'Neill, M. J.; Gray, Allan P.; "Use of magnetic transitions in temperature calibration and performance evaluation of thermogravimetric systems," *Thermochimica Acta* (1970), 1(1), 29-38.

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Stanley J. Pruchnic, Jr., whose telephone number is **(571) 272-2248**. The examiner can normally be reached on weekdays (Monday through Friday) from 7:30 AM to 4:00 PM. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Diego F. F. Gutierrez can be reached at **(571) 272-2245**.


The **Official FAX** number for Technology Center 2800 is **(703) 872-9306** for **all official communications**.

Any inquiry of a general nature or relating to the status of this application or proceeding may be directed to the official USPTO website at <http://www.uspto.gov/> or you may call the **USPTO Call Center** at **800-786-9199** or 703-308-4357. The Technology Center 2800 Customer Service FAX phone number is (703) 872-9317.

The cited U.S. patents and patent application publications are available for download via the Office's PAIR. As an alternate source, all U.S. patents and patent application publications are available on the USPTO web site ([www.uspto.gov](http://www.uspto.gov)), from the Office of Public Records and from commercial sources.

Private PAIR provides external customers Internet-based access to patent application status and history information as well as the ability to view the scanned images of each customer's own application file folder(s).

For inquiries relating to Patent e-business products and service applications, you may call the **Patent Electronic Business Center (EBC)** at **703-305-3028** or toll free at **866-217-9197** between the hours of **6 a.m. and midnight Monday through Friday EST**, or by e-mail at: [ebc@uspto.gov](mailto:ebc@uspto.gov). Additional information is available on the Patent EBC Web site at: <http://www.uspto.gov/ebc/index.html>.



**DIEGO F. F. GUTIERREZ**  
**SUPERVISORY PATENT EXAMINER**  
**TECHNOLOGY CENTER 2800**



**Stanley J. Pruchnic, Jr.**  
**11/22/04**

**CHRISTOPHER W. FULTON**  
**PRIMARY EXAMINER**